

## CERTIFICATE OF FACTORY CALIBRATION

---

*We hereby certify that all of our devices are calibrated at the factory and meet all specifications listed in our documentation.*

Device under Test:	DC Electronic Load
Model:	CH9720CU
Serial - Number:	L01316084
Calibration Date:	12.Mar.2021
Environment Temperature:	23.0°C ±0,5°C
Humidity:	35% ± 10% RH
Calibration Equipment:	Fluke 8845A
Calibrated by:	Penglin Cheng

---

Changzhou Beich Electronic Technology Co., Ltd

☎ 86-519-85503555

☎ 86-519-85503113

✉ [sales@beich.com.cn](mailto:sales@beich.com.cn)

🌐 [www.beich.com.cn](http://www.beich.com.cn)

Temperature: 23°C ± 5°C  
Calibration Equipment Used: Fluke 8845A

Relative Humidity: < 70%

一: Voltage Measurement

Voltage (V)	Actual value (V)	Standard value (V)	Allowable range	Pass/Fail
0~36	0.2004	0.2000	0.1944~0.2056	Pass
	1.0008	1.0000	0.9936~1.0064	Pass
	5.0005	5.0000	4.9896~5.0104	Pass
	10.003	10.000	9.9846~10.0154	Pass
	15.004	15.000	14.9796~15.0204	Pass
	25.002	25.000	24.9696~25.0304	Pass
	35.006	35.000	34.9596~35.0404	Pass
36~360	50.007	50.000	49.914~50.086	Pass
	100.05	100.00	99.864~100.136	Pass
	200.03	200.00	199.692~200.308	Pass
	300.06	300.00	299.592~300.408	Pass

二: Current Measurement

Current (A)	Actual value (A)	Standard value (A)	Allowable range	Pass/Fail
0~3.000	0.1003	0.1000	0.0969~0.1031	Pass
	0.9998	1.0000	0.996~1.004	Pass
	1.9996	2.0000	1.995~2.005	Pass
	2.9994	3.0000	2.994~3.006	Pass
3.00~30.00	3.0005	3.0000	2.949~3.051	Pass
	10.002	10.000	9.935~10.065	Pass
	20.003	20.000	19.915~20.085	Pass
	30.006	30.000	29.895~30.105	Pass

Conclusion: Pass

Identifier: \_\_\_\_\_